

✓ **REFERENCE**

5 calibrators KIT	MPREK-000	5 x 1 ml	2-8°C
Human multiparametric biological fluid standardised from the reference ERM-DA470k/IFCC, sodium azide (< 1g/l)			
Lot #	17H28		
Expiry date	01-2019		
Control date	03-10-17		
Quality control report #	DGM-QAC-REP-17139		
Document prepared and signed by	L. Ginneberge		

✓ **SAMPLES AND REFERENCE VALUES**

See the corresponding reagents technical sheet.

✓ **COMPOSITION**

Multiparametric calibrators are human biological fluids diluted in HEPES pH 7.4 buffer containing stabilisers, sodium azide (<1g/l) as preservative and the following human proteins:

albumin, alpha-1 antitrypsin, alpha-1 glycoprotein acid, alpha-2 macroglobulin, antithrombin III, complement C3, complement C4, ceruloplasmin, haptoglobin, IgA, IgG, IgM, prealbumin and transferrin.

✓ **PRINCIPLE OF TEST**

The human proteins of calibrators react upon a specific antibody for corresponding protein and the turbidity induced by the formation of immune complexes is recorded at appropriate wavelength. The turbidity measured is directly proportional to the antigen concentration of the calibrators which can be used for the quantitative determination of this antigen in immunoturbidimetry.

✓ **PRECAUTIONS**

For in vitro single diagnostic use. To be handled by entitled Personnel. Products from human source were tested and found free from HBsAg and antibodies to HCV and HIV but this material should be treated just as carefully as potentially infective. Products containing sodium azide have to be handled with care; avoid ingestion and contact with skin and mucous membranes. Sodium azide may react with lead or copper plumbing to form highly explosive metal azides.

✓ **ANALYTICAL PERFORMANCES**

See the corresponding reagents technical sheet.

✓ **PREPARATION AND REAGENTS STABILITY**

The calibrators are ready for use; once opened, they are stable until expiry date if stored stoppered in appropriate temperature conditions and without any contamination (avoid pipetting and decantation).

✓ **METHOD OF ANALYSIS AND CALCULATION**

See the corresponding reagents technical sheet.

✓ **QUALITY CONTROL**

Accuracy and reproducibility: analytical performances can be checked with the internal quality control serum of the laboratory or with the Liquichek™ (BIO-RAD) Control sera (see the values range obtained with DiAgam reagents and indicated on the accompanying BIO-RAD sheet).

Calibration: calibration curve and stability of calibration curve can be validated with the DiAgam calibration control (MPCON-002, MPCOS-002 and MPCOX-002).

In case of analytical performances modification, calibrate the method again and contact the manufacturer if modifications are subsisting.

✓ **BIBLIOGRAPHY**

- (1) Certification of proteins in the human serum. Certified Referenced Material ERM®-DA470k/IFCC. I. Zegers et al. <http://irmm.jrc.ec.europa.eu/>
- (2) S. Blirup-Jensen et al. protein standardization V: value transfer. A practical protocol for the assignment of serum protein values from a reference material to a target material. Clin Chem Lab Med (2008); 46(10): 1470-1479.
- (3) G. Merlini et al. Standardizing plasma protein measurements worldwide: a challenging enterprise. Clin Chem Lab Med (2010); 48(11): 1567-1575.



MPREKFTEN 03-10-17 v06 rev.02

Proteins:	CAL 1		CAL 2		CAL 3		CAL 4		CAL 5	
	g/l		g/l		g/l		g/l		g/l	
	certified val.	U*	certified val.	U*	certified val.	U*	certified val.	U*	certified val.	U*
Albumin	6,77	0,339	13,22	0,661	25,63	1,282	50,84	2,542	105,17	5,259
Alpha-1-Antitrypsin **	0,19	0,010	0,37	0,019	0,74	0,037	1,50	0,075	2,99	0,150
Alpha-1-Acid Glycoprotein	0,12	0,006	0,24	0,012	0,47	0,024	0,96	0,048	1,90	0,095
Alpha-2-Macroglobulin	0,35	0,018	0,66	0,033	1,27	0,064	2,53	0,127	5,06	0,253
Antithrombin III*	0,054	0,003	0,070	0,004	0,138	0,007	0,297	0,015	0,465	0,023
Complement C3	0,22	0,011	0,44	0,022	0,88	0,044	1,78	0,089	3,56	0,178
Complement C4	0,040	0,002	0,080	0,004	0,158	0,008	0,319	0,016	0,639	0,032
Ceruloplasmin*	0,083 ***	0,004	0,162 ***	0,008	0,303 ***	0,015	0,607 ***	0,030	1,186 ***	0,059
Haptoglobin	0,19	0,010	0,39	0,020	0,76	0,038	1,52	0,076	3,07	0,154
IgA	0,31	0,016	0,61	0,031	1,22	0,061	2,48	0,124	4,95	0,248
IgG	1,55	0,078	3,13	0,157	6,27	0,314	12,63	0,632	25,27	1,264
IgM	0,14	0,007	0,28	0,014	0,57	0,029	1,15	0,058	2,30	0,115
Prealbumin	0,033	0,002	0,066	0,003	0,127	0,006	0,256	0,013	0,512	0,026
Transferrin	0,40	0,020	0,80	0,040	1,58	0,079	3,20	0,160	6,40	0,320

U* : The certified uncertainty is the half-width of the 95 % confidence interval of the mean.

Values assigned from the reference ERM-DA470k/IFCC.

*AT-III and Ceruloplasmin is referenced to external controls.

** Values compatible with the new Alpha-1-Antitrypsin reagent and incompatible with the old reagent. In case of doubt, contact the manufacturer.

*** Values compatible with the new Ceruloplasmin reagent and incompatible with the old reagent. In case of doubt, contact the manufacturer.